


**SGS**

**UPSTREAM AND DOWNSTREAM ATMOSPHERIC EMISSIONS  
OF THE FILTRATION SYSTEM “PROTOTYPE WITH NPE TECHNOLOGY”**

PERFORMED FOR:  
**NEW POWER ENERGY KFT**



**SGS** is the world's leading inspection, verification, testing and certification company. Recognised as the global benchmark for quality and integrity, We provide innovative services and **solutions** for every part of the environmental industry. Our global network of offices and laboratories, alongside our dedicated team, allows us to respond to your needs, when and where they occur.

**UPSTREAM AND DOWNSTREAM  
ATMOSPHERIC EMISSIONS OF THE  
FILTRATION SYSTEM “PROTOTYPE WITH NPE  
TECHNOLOGY”**

**RT 6096 - 2022**

FEBRUARY 28, 2022

Prepared by

**SGS ITALIA S.P.A.**  
**ENVIRONMENTAL SERVICES**  
VIA CAMPODORO, 25  
35010 VILLAFRANCA PADOVANA – PD

Performed for

**NEW POWER ENERGY KFT**  
KASSAI UTCA, 9  
1043 – BUDAPEST ()



## INDEX

---

1. INTRODUCTION.....	4
2 DESCRIPTION OF THE SAMPLING CARRIED OUT .....	5
3 ANALYTICAL RESULTS.....	7

---

This Relation is issued by the Company in accordance with the SGS General Conditions for Inspection and Control Services (at <http://www.sgs.com/en/Terms-and-Conditions.aspx>). The release of this Relation does not exempt the negotiating parties from exercising the rights and from fulfilling the obligations deriving from the agreement stipulated between them. Any contrary agreement is not enforceable by the Company. The liability of the Company under this Relation is limited to the case of proven gross negligence and in any case to an amount not exceeding ten times the fees and commissions due.



## 1. INTRODUCTION

---

We hereby send you the summary of the results of the sampling and measurements of the atmospheric emissions carried out on 25<sup>TH</sup> of February 2022 at your "Prototype With NPE Technology" Filtration System

The results contained in the test reports no. PD22-00604, and the results of the processing refer exclusively to the operating conditions in place in the period in which this survey was carried out.

This technical report must not be reproduced except in full.

---

### Allegati:

---

Annex A: Calibration certificates of the instrumentation used;

Annex B: Analytical Report PD22-00604.

---



## 2 DESCRIPTION OF THE SAMPLING CARRIED OUT

On February 25, 2022, on direct assignment by the client company NEW POWER ENERGY KFT, located in Budapest (H), **SGS ITALIA SPA** technicians performed a series of environmental monitoring of the emissions produced by a combustion generator powered by LPG, with a power of 8 kW, at the Client's test laboratory, located in Via Viazza, 94 / D in San Prospero (MO) Italy, at 4,800 rpm, upstream of the filtration system called "prototype with NPE technology", in order to check the concentrations of some pollutants emitted into the atmosphere during the combustion of the generator.

In particular, it should be noted that the sampling performed has the purpose of verifying the concentrations of pollutants (indicated below) at the outlet of the two sampling points, declared by the client company as "Upstream and Downstream" of an air filtration system called "Prototype with NPE technology".

Below is an image of the prototype (external box) on which the upstream and downstream self-checks were carried out.





Below is the list of parameters monitored in the 2 upstream and downstream points of your System:

Parameter	Method
Oxygen (O <sub>2</sub> )	UNI EN 14789:2017
Carbon dioxide (CO <sub>2</sub> )	EPA 3A 2017
Nitrogen oxides (NO <sub>x</sub> )	UNI EN 14792:2017
Sulphur dioxide (SO <sub>2</sub> )	UNI CEN/TS 17021/2017
Total Organic Carbon (TOC)	UNI EN 12619:2013
PM10-PM2,5	UNI EN ISO 23210:2009



### 3 ANALYTICAL RESULTS

Below is a summary of the final data of the samplings performed:

parameter	before NPE reactor (Upstream)	after NPE reactor (Downstream)
	Concentration measured	Concentration measured
<b>Reference Analytical Report n: PD22-00604</b>		
Flow rate of humid fumes	47±1.00 Nm <sup>3</sup> /h	41±1.00 Nm <sup>3</sup> /h
Flow rate of dry fumes	36±2.00 Nm <sup>3</sup> /h	41±2.00 Nm <sup>3</sup> /h
Water vapor	23.90±2.30 %vv	<2.0 %vv
Oxygen (O <sub>2</sub> )	1.89±0.12 %vv	20.9±1.5 %vv
Carbon dioxide (CO <sub>2</sub> )	12.33±0.39 %vv	0.01486±0.00149 %vv
Carbon dioxide (CO <sub>2</sub> )	123300±12330 ppm	148.6±14.86 ppm (*)
Sulphur dioxide (SO <sub>2</sub> )	<2.9 mg/Nm <sup>3</sup>	<2.9 mg/Nm <sup>3</sup>
Nitrogen oxides (NO <sub>x</sub> )	5400±460 mg/Nm <sup>3</sup>	<2.1 mg/Nm <sup>3</sup>
Volatile organic compounds (VOCs) expressed as Total Organic Carbon	9.03±0.90 mg/Nm <sup>3</sup>	1.73±0.90 mg/Nm <sup>3</sup>
Carbon monoxide (CO)	<1.3 mg/Nm <sup>3</sup>	<1.3 mg/Nm <sup>3</sup>
PM2.5	0.082±0.023 mg/Nm <sup>3</sup>	<0.043 mg/Nm <sup>3</sup>
PM10	0.341±0.097 mg/Nm <sup>3</sup>	0.064±0.018 mg/Nm <sup>3</sup>

(\*) measurement performed with TSI 7545 IAQ-Calc

Preparato da	Firmata da
Enrico Costa (Project Leader) 	Dr. Gianluca Martinati Head of Laboratory Interprovincial Order of Chemists and the Physicists of Veneto – Padova/n°787/24 





## **Annex n. 6096/A**

**TÜV RHEINLAND  
ENERGIE UND UMWELT GMBH**



Report on the test according to the reference standard EN 15058 of the measuring system PG-350E for the measured component CO of the company Horiba Europe GmbH, Leichlingen

**TÜV-Report-No.: 936/21221241/A**  
Cologne, February 26, 2013

[www.umwelt-tuv.de](http://www.umwelt-tuv.de)



[luft@de.tuv.com](mailto:luft@de.tuv.com)

**The department of Environmental Protection of TÜV Rheinland Energie und Umwelt GmbH**  
is accredited for the following work areas:

- Determination of air quality and emissions of air pollution and odour substances;
- Inspection of correct installation, function and calibration of continuously operating emission measuring instruments, including data evaluation and remote emission monitoring systems;
- Combustion chamber measurements;
- Performance testing of measuring systems for continuous monitoring of emissions and ambient air, and of electronic data evaluation and remote emission monitoring systems;
- Determination of stack height and air quality projections for hazardous and odour substances;
- Determination of noise and vibration emissions and pollution, determination of sound power levels and execution of sound measurements at wind energy plants

**according to EN ISO/IEC 17025.**

The accreditation is valid up to 22-01-2018. DAkkS-register number: D-PL-11120-02-00.

The publication of extracts is subject to approval by.

**TÜV Rheinland Energie und Umwelt GmbH**  
**D - 51105 Cologne, Am Grauen Stein, Tel: 0221 806-2756, Fax: 0221 806-1349**

Report on the test according to the reference standard EN 15058 of the measuring system PG-350E for the measured component CO of the company Horiba Europe GmbH, Leichlingen, Report-No: 936/21221241/A

**Blank page**

Report on the test according to the reference standard EN 15058 of the measuring system PG-350E for the measured component CO of the company Horiba Europe GmbH, Leichlingen, Report-No.: 936/21221241/A

Page 3 of 73

## OVERVIEW

The company Horiba Europe GmbH has commissioned TÜV Rheinland Energie und Umwelt GmbH with the performance of a test of the PG-350E measuring system for the measured component CO according to the European reference standard EN 15058.

The measuring system is an extractive NDIR sensor, which is integrated in the portable Horiba PG-350E analyser. The instrument tested is able to measure in addition to CO components NOx, SO<sub>2</sub>, CO<sub>2</sub> and O<sub>2</sub>. The report in hand is however concerned solely with the measurement of CO in accordance with EN 15058. To allow the measurement of NO<sub>x</sub> at plants with a NO<sub>2</sub> / NO<sub>x</sub> ratio higher than 10 %, the system can be equipped in accordance with standard EN 14792 for nitrogen oxides with a Horiba permeation dryer PD-100.

The test work took place in the laboratory. The tested measuring range was:


Component	Measuring range		
Carbon monoxide	CO	0 - 75	mg/m <sup>3</sup>

Furthermore an exemplary field test in the clean gas of a waste incineration plant followed the laboratory test.

The requirements on the performance criteria and on the overall uncertainty of the method according to the standard EN 15058 have been fulfilled during the test.

Therefore the measuring system PG-350E with the extractive NDIR sensor is suitable for the use as a standard reference method for the measurement of carbon monoxide in waste gas according to the standard EN 15058.

Department of Environmental Protection



Dipl.-Ing. Guido Baum



Dr. rer. nat. Peter Wilbring

Cologne, February 26, 2013

936/21221241/A

Report on the test according to the reference standard EN 15058 of the measuring system PG-350E for the measured component CO of the company Horiba Europe GmbH, Leichlingen, Report-No: 936/21221241/A

**Blank page**

**TÜV RHEINLAND  
ENERGIE UND UMWELT GMBH**



Report on the test according to the reference standard EN 14792 of the measuring system PG-350E for the measured component NO<sub>x</sub> of the company Horiba Europe GmbH, Leichlingen

**TÜV-Report-No.: 936/21221241/B**  
Cologne, February 26, 2013

[www.umwelt-tuv.de](http://www.umwelt-tuv.de)



[luft@de.tuv.com](mailto:luft@de.tuv.com)

**The department of Environmental Protection of TÜV Rheinland Energie und Umwelt GmbH**  
is accredited for the following work areas:

- Determination of air quality and emissions of air pollution and odour substances;
- Inspection of correct installation, function and calibration of continuously operating emission measuring instruments, including data evaluation and remote emission monitoring systems;
- Combustion chamber measurements;
- Performance testing of measuring systems for continuous monitoring of emissions and ambient air, and of electronic data evaluation and remote emission monitoring systems;
- Determination of stack height and air quality projections for hazardous and odour substances;
- Determination of noise and vibration emissions and pollution, determination of sound power levels and execution of sound measurements at wind energy plants

**according to EN ISO/IEC 17025.**

The accreditation is valid up to 22-01-2018. DAkkS-register number: D-PL-11120-02-00.

The publication of extracts is subject to approval by  
**TÜV Rheinland Energie und Umwelt GmbH**  
**D - 51105 Köln, Am Grauen Stein, Tel: 0221/806-2756, Fax: 0221/806-1349**

Report on the test according to the reference standard EN 14792 of the measuring system PG-350E for the measured component NOx of the company Horiba Europe GmbH, Leichlingen, Report-No: 936/21221241/B

Blank page

## OVERVIEW

The company Horiba Europe GmbH has commissioned TÜV Rheinland Energie und Umwelt GmbH with the performance of a test of the PG-350E measuring system for the measured component NOx according to the European reference standard EN 14792

The measuring system is an extractive chemiluminescence sensor, which is integrated in the portable Horiba PG-350E analyser. The instrument tested is able to measure in addition to NO<sub>x</sub> components CO, SO<sub>2</sub>, CO<sub>2</sub> and O<sub>2</sub>. The report in hand is however concerned solely with the measurement of CO in accordance with EN 14792. To allow the measurement of NO<sub>x</sub> at plants with a NO<sub>2</sub> / NO<sub>x</sub> ratio higher than 10 %, the system can be equipped in accordance with standard EN 14792 for nitrogen oxides with a Horiba permeation dryer PD-100.

The test work took place in the laboratory. The tested measuring range was:

Component	Measuring range	
Nitrogen oxides	NO <sub>x</sub>	0 - 205 mg/m <sup>3</sup> as NO <sub>2</sub>
	NO <sub>x</sub>	0 - 134 mg/m <sup>3</sup> as NO

Furthermore, an exemplary field test in the clean gas of a waste incineration plant followed the laboratory test.

The requirements on the performance criteria and on the overall uncertainty of the method according to the standard EN 14792 have been fulfilled during the test.

Therefore the measuring system PG-350E with the chemiluminescence detector is suitable for the use as a standard reference method for the measurement of nitrogen oxides in waste gas according to the standard EN 14792.

Department of Environmental Protection

Guido Baum

Dipl.-Ing. Guido Baum

Peter Wilbring

Dr. rer. nat. Peter Wilbring

Cologne, February 26, 2013  
936/21221241/B

Report on the test according to the reference standard EN 14792 of the measuring system PG-350E for the measured component NOx of the company Horiba Europe GmbH, Leichlingen, Report-No: 936/21221241/B

**Blank page**

**TÜV RHEINLAND  
ENERGIE UND UMWELT GMBH**



Report on the test according to the reference standard EN 14789 of the measuring system PG-350E for the measured component O<sub>2</sub> of the company Horiba Europe GmbH, Leichlingen

**TÜV-Report-No.: 936/21221241/C**  
Cologne, February 26, 2013

[www.umwelt-tuv.de](http://www.umwelt-tuv.de)



[luft@de.tuv.com](mailto:luft@de.tuv.com)

**The department of Environmental Protection of TÜV Rheinland Energie und Umwelt GmbH**  
is accredited for the following work areas:

- Determination of air quality and emissions of air pollution and odour substances;
- Inspection of correct installation, function and calibration of continuously operating emission measuring instruments, including data evaluation and remote emission monitoring systems;
- Combustion chamber measurements;
- Performance testing of measuring systems for continuous monitoring of emissions and ambient air, and of electronic data evaluation and remote emission monitoring systems;
- Determination of stack height and air quality projections for hazardous and odour substances;
- Determination of noise and vibration emissions and pollution, determination of sound power levels and execution of sound measurements at wind energy plants

**according to EN ISO/IEC 17025.**

The accreditation is valid up to 22-01-2018. DAkkS-register number: D-PL-11120-02-00.

The publication of extracts is subject to approval by.

**TÜV Rheinland Energie und Umwelt GmbH**  
**D - 51105 Cologne, Am Grauen Stein, Tel: 0221 806-2756, Fax: 0221 806-1349**

Report on the test according to the reference standard EN 14789 of the measuring  
system PG-350E for the measured component O2 of the company Horiba Europe  
GmbH, Lelchingen, Report-No: 936/21221241/C

**Blank page**



# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

## **PG-350E Multi-component Gas Analyser**

Manufactured by:

**Horiba Europe GmbH**

Julius Kronenberg Straße 9  
42799 Leichlingen  
Germany

Has been assessed by Sira Certification Service  
And for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Emission  
Monitoring Systems, Version 3.4 dated July 2012, Annex F; Transportable Systems,  
EN15267-3:2007,  
& QAL 1 as defined in EN 14181: 2004**

Certification Ranges:

CO	0 to 75 mg/m <sup>3</sup> , 0 to 6250mg/m <sup>3</sup>
CO <sub>2</sub>	0 to 20 Vol.%
NO <sub>x</sub>	0 to 134 mg/m <sup>3</sup>
O <sub>2</sub>	0 to 25 Vol.%, * 0 to 10Vol.%
SO <sub>2</sub>	0 to 143 mg/m <sup>3</sup> , 0 to 8580mg/m <sup>3</sup>

**\*(Additional testing for these gases has been conducted for certification to Annex F)**

Project No. : 16A29871  
Certificate No : Sira MC130223/01  
Initial Certification : 28 February 2013  
This Certificate issued : 13 January 2014  
Renewal Date : 27 February 2018

R Cooper I Eng MInst MC  
Technical Director

MCERTS is operated on behalf of the Environment Agency by

**Sira Certification Service**

12 Acorn Industrial Park, Crayford Road, Crayford  
Dartford, Kent, UK DA1 4AL  
Tel: +44 (0)1322 520500 Fax: +44 (0)1322 520501

*This certificate may only be reproduced in its entirety and without change*

*Registered Office: Rake Lane, Eccleston, Chester, UK CH4 9JN*

*To authenticate the validity of this certificate please visit [www.siracertification.com/mcerts](http://www.siracertification.com/mcerts)*





---

## Approved Site Application

*Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at [www.mcerts.net](http://www.mcerts.net)*

On the basis of the assessment and the ranges required for compliance with EU Directives, this instrument is considered suitable for use as an SRM and for verifying and calibrating installed CEMS, according to the requirements of EN14181. This portable analyser is also considered suitable for use as a back-up CEM, excluding the measurement of daily mean SO<sub>2</sub> values for plants that operate within the scope of the 2000/76/EC (WID) Directive.

The field test was conducted on a municipal waste incinerator.

### Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

- TÜV report 936/21217617/A\_en\_draft dated 5<sup>th</sup> October 2012
- TUV report 936/20130327 dated 27<sup>th</sup> March 2013
- TUV report 936/21221241/A dated 26<sup>th</sup> February 2013 (SRM data for CO)
- TUV report 936/21221241/B dated 26<sup>th</sup> February 2013 (SRM data for NO<sub>x</sub>)
- TUV report 936/21221241/C dated 26<sup>th</sup> February 2013 (SRM data for O<sub>2</sub>)

Certificate No : Sira MC130223/01  
This Certificate issued : 13 January 2014

*This certificate may only be reproduced in its entirety and without change.  
To authenticate the validity of this certificate please visit [www.siracertification.com/mcerts](http://www.siracertification.com/mcerts)*

**EG-Konformitätserklärung**  
EC Declaration of Conformity

**Die Firma**  
The company

Mess- & Analysetechnik GmbH  
D-51381 Leverkusen  
Benzstr. 23 - 25  
Germany



erklärt hiermit, dass folgende Geräte  
Declares hereby, that the following devices

Geräteart Device	Kohlenwasserstoff Analysator (FID) Hydrocarbon Analyser (FID)
Typbezeichnung Type	Thermo-FID ES, Thermo-FID TG, Thermo-FID PT, Thermo-FID KA Thermo-FID FE, Thermo-FID-MSU, Thermo-FID MK

**Mit den Vorschriften folgender Europäischer Richtlinien übereinstimmen**  
Complies with the requirements of the following European directives

89/336/EWG 89/336/ECC	EMV Richtlinie EMC Directive
73/23/EWG 73/23/ECC	Niederspannungsrichtlinie Low Voltage Directive
97/23/EG 97/23/EC	Druckgeräterichtlinie Pressure Equipment Directive

Wir beschreinigt hiermit, dass Konstruktion, Herstellung und Prüfung dieser Geräte den Anforderungen der Druckgeräterichtlinie entsprechen. Die Ausrüstungssteile fallen unter Artikel 3, Absatz (3) der DGR 97/23/EG. Die CE Kennzeichnung bezieht sich daher nicht auf diese Richtlinie.

We declare herewith that design, production and check-out of the devices are in conformity with the Pressure Equipment Directive. The peripheral accessories are subject to Article 3, Paragraph (3) of PED 97/23/EC. Therefore the CE marking does not refer to this directive.

**Angewandte harmonisierte Normen**  
Harmonized standards used

EN61326-1:1997 + A1:1998	Elektrische Betriebsmittel für Leittechnik und Laboreinsatz – EMV Anforderungen Teil 1: Allgemeine Anforderungen Electrical equipment for measurement, control and laboratory use - EMC requirements Part 1: General requirements
EN61010-1:1993 + A1:1992	Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte Teil 1: Allgemeine Anforderungen Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements

Leverkusen, 17.01.2005

(Unterschrift / Signature)  
**Birgit Kühn**  
(Name)  
Geschäftsführerin / General Manager  
(Funktion / Function name)

Diese Erklärung beschreinigt die Übereinstimmung mit den genannten Richtlinien, beinhaltet jedoch keine Zusicherung von Eigenschaften.  
Die Sicherheitshinweise und Installationsvorschriften sind zu beachten.  
This declaration confirms the compliance with the announced directives but does not include the assurance of properties.  
The safety and installation instructions of the documentation have to be followed.

**Carbagas**

*Gaz sur mesure.*



AR-332

Certificat d'analyse

N° 2600963

N° de commande client : 4509470263

Stabilité garantie jusqu'au : 01.03.2022

N° de mélange :	31635
Classe de réalisation :	Saphir
N° de commande interne :	1299047542
N° client :	A360003_IY
Nom du client :	Air Liquide Italia Service S.r.l.
Adresse du client :	S.P 14 Rivoltana Km 6, 20090 RODANO
Code Produit :	G000003014A001

Composant	Qualité	Concentration Demandée	Concentration Réalisée	Incertitude (%)
		(mol.)		
NO	30	100.0 ppm	99.98	2.00%
CO	47	100.0 ppm	100.31	2.00%
SO2	30	100.0 ppm	103.84	2.00%
CO2	50	5.00 %	5.04	2.00%
N2	60	Reste		

Type d'emballage :	10.0 l. Aluminium
N° de bouteille :	AEGPWAS
Pression finale :	150 Bar
Raccord vanne :	Raccord client inox
Pression min. d'utilisation :	2 Bar
Temp. Minimum :	-10 °C

Date et lieu de l'analyse : Domdidier, 31.03.2020

Il s'agit d'un certificat réalisé par ordinateur, qui est valable sans signature.  
CARBAGAS AG  
route d'Avenches 89 1569 Domdidier  
Visa : C. Schenk





Air Liquide Italia Service S.r.l.  
**Laboratorio Specialty Gases**  
 20090 Rodano (MI) - S.P. 14 Rivoltana km 6  
 Tel. 02 95757 244/225 - Fax 02 95320616  
 Industria.airliquide.it

## CERTIFICATO



Cliente	SGS Italia	Data	19/02/2020
Richiedente	UO Padova	Protocollo	2020-565
Recipiente	10 LT	Natura del contenuto	Miscela
Barcode	ADRR6C8	Nr. Scheda Mix	2367

COMPONENTE	Concentrazione			Incertezza Espansa (**)
	Nominale	Tolleranza	Valore misurato	
Propano C3H8	100 ppm	± 5 %	97.80 ppm	± 2 %
Ossigeno O2	20 %	± 5 %	19.907 %	± 2 %

AR 318

Complemento	Azoto	Concentrazione	MOL.
Temperatura min. di utilizzo	5 °C	Pressione di riempimento	151 bar
Scadenza miscela (Mesi)	36		
Volume di gas a 15°C 1013,25 mbar	1491 Litri	Pressione min. di utilizzo	5 bar
Normativa di riferimento per la preparazione: ISO 6142 Normativa di riferimento per analisi: ISO 6143 La miscela è stata preparata con il metodo gravimetrico su bilance tarate con masse certificate da Centro di Taratura LAT N° 055.			

(\*\*) Intervallo di confidenza 95%


AIR LIQUIDE ITALIA Service S.r.l.

L'Analista

Eleonora Gurrieri

## Carta di controllo

Miscela: N2 Val. atteso: 0,0 Unità di misura: ppm  
Mat. di Rif.: bombola N2 5.0 Metodo di prova: controllo zero Lim. Inf. -0,2 Lim. Sup.: 0,2  
Strumento di analisi TSI Q-Trak 8554



Valore  
Data  
Tecnico  
Note

0,0	0,0	0,0
10/05/20	10/05/21	24/02/22
Fabris	Baccarin	Baccarin

## Carta di controllo

Miscela:

CO2 50400 ppm

Val. atteso: 5040,0

Unita' di misura:

ppm

Mat. di Rif.:

AR 332 miscela

Lim. Inf


4939,2

Strumento di analisi

multiparametrica TSI Q-Trak 8554

Lim. Sup.:

5140,8



Valore  
Data  
Tecnico  
Note

5040,0	5044,0	5038,0
10/05/20	10/05/21	24/02/22
Fabris	Baccarin	Baccarin

Diluizione 1:10 con Diluitore 2 (MFC 12-MFC 13)

**Carbagas**

*Gaz sur mesure.*



AR-332

Certificat d'analyse

N° 2000963

N° de commande client : 4509470263

Stabilité garantie jusqu'au : 01.03.2022

N° de mélange :	31635
Classe de réalisation :	Saphir
N° de commande interne :	1299047542
N° client :	A360003_IY
Nom du client :	Air Liquide Italia Service S.r.l.
Adresse du client :	S.P 14 Rivoltana Km 6, 20090 RODANO
Code Produit :	G000003014A001

Composant	Qualité	Concentration Demandée (mol.)	Concentration Réalisée	Incertitude (%)
NO	30	100.0 ppm	99.98	2.00%
CO	47	100.0 ppm	100.31	2.00%
SO2	30	100.0 ppm	103.84	2.00%
CO2	50	5.00 %	5.04	2.00%
N2	60	Reste		

Type d'emballage :	10.0 l. Aluminium
N° de bouteille :	AEGPWAS
Pression finale :	150 Bar
Raccord vanne :	Raccord client inox
Pression min. d'utilisation :	2 Bar
Temp. Minimum :	-10 °C

Date et lieu de l'analyse : Domdidier, 31.03.2020

Il s'agit d'un certificat réalisé par ordinateur, qui est valable sans signature.

CARBAGAS AG

route d'Avenches 89 1569 Domdidier

Visa : C. Schenk





Centro di Taratura LAT 159  
Calibration Centre

Laboratorio Accreditato di Taratura  
Accredited Calibration Laboratory



ACCREDIA  
ENTE ITALIANO DI ACCREDITAMENTO

LAT N° 159 T

SGS Italia SpA  
Environmental Services  
Via Campodoro, 25  
35010 Villafranca Padovana PD  
t +39 049 9050013  
f +39 049 9050065  
e [sgs.eco@sgs.com](mailto:sgs.eco@sgs.com)

Pagina 1 di 4  
Page 1 of 4

CERTIFICATO DI TARATURA LAT 159 1542  
Certificate of Calibration

- data di emissione <i>date of issue</i>	10 febbraio 2021
- cliente <i>customer</i>	SGS ITALIA SPA Via Campodoro, 25 35010 – Villafranca P.na - PD
- destinatario <i>receiver</i>	---

Il presente certificato di taratura è emesso in base all'accreditamento LAT 159 rilasciato in accordo ai decreti attuativi della legge n. 273/1991 che ha istituito il Sistema Nazionale di Taratura (SNT). ACCREDIA attesta le capacità di misura e di taratura, le competenze metrologiche del Centro e la riferibilità delle tarature eseguite ai campioni nazionali e internazionali delle unità di misura del Sistema Internazionale delle Unità (SI). Questo certificato non può essere riprodotto in modo parziale, salvo espressa autorizzazione scritta da parte del Centro.

<u>Si riferisce a</u> <i>Referring to</i>	
- oggetto <i>item</i>	Flussimetro massico – Controllo strumento
- costruttore <i>manufacturer</i>	MKS - MKS
- modello <i>model</i>	GE50A013503SMV020 – PR4000BF2V2
- matricola <i>serial number</i>	022239892 – 018957352
- data di ricevimento oggetto <i>date of receipt of item</i>	25 gennaio 2021
- data delle misure <i>date of measurements</i>	9 febbraio 2021
- registro di laboratorio <i>laboratory reference</i>	R20-28

*This certificate of calibration is issued in compliance with the accreditation LAT 159 granted according to decrees connected with Italian law No. 273/1991 which has established the National Calibration System. ACCREDIA attests the calibration and measurement capability, metrological competence of the Centre and the traceability of calibration results to the national and international standards of the International System of Units (SI).*  
*This certificate may not be partially reproduced, except with the prior written permission of the issuing Centre.*

I risultati di misura riportati nel presente Certificato sono stati ottenuti applicando le procedure di taratura citate alla pagina seguente, dove sono specificati anche i campioni o gli strumenti che garantiscono la catena di riferibilità del Centro e i rispettivi certificati di taratura in corso di validità. Essi si riferiscono esclusivamente all'oggetto in taratura e sono validi nel momento e nelle condizioni di taratura, salvo diversamente specificato.

*The measurement results reported in this Certificate were obtained following the calibration procedures given in the following page, where the reference standards or instruments are indicated which guarantee the traceability chain of the laboratory, and the related calibration certificates in the course of validity are indicated as well. They relate only to the calibrated item and they are valid for the time and conditions of calibration, unless otherwise specified.*

Le incertezze di misura dichiarate in questo documento sono state determinate conformemente alla Guida ISO/IEC 98 e al documento EA-4/02. Solitamente sono espresse come incertezza estesa ottenuta moltiplicando l'incertezza tipo per il fattore di copertura  $k$  corrispondente ad un livello di fiducia di circa il 95 %. Normalmente tale fattore  $k$  vale 2.

*The measurement uncertainties stated in this document have been determined according to the ISO/IEC Guide 98 and to EA-4/02. Usually, they have been estimated as expanded uncertainty obtained multiplying the standard uncertainty by the coverage factor  $k$  corresponding to a confidence level of about 95%. Normally, this factor  $k$  is 2.*

Direzione tecnica  
(Approving Officer)

*Luca Franchi*



Centro di Taratura LAT 159  
Calibration Centre

Laboratorio Accreditato di Taratura  
Accredited Calibration Laboratory



ACCREDIA  
CENTRO ITALIANO DI ACCREDITAMENTO

LAT N° 159 T

SGS Italia SpA  
Environmental Services  
Via Campodoro, 25  
35010 Villafranca Padovana PD  
t +39 049 9050013  
f +39 049 9050065  
e [sgs.eco@sgs.com](mailto:sgs.eco@sgs.com)

Pagina 1 di 4  
Page 1 of 4

CERTIFICATO DI TARATURA LAT 159 1541  
Certificate of Calibration

- data di emissione <i>date of issue</i>	10 febbraio 2021
- cliente <i>customer</i>	SGS ITALIA SPA Via Campodoro, 25 35010 – Villafranca P.na - PD
- destinatario <i>receiver</i>	-----

Il presente certificato di taratura è emesso in base all'accreditamento LAT 159 rilasciato in accordo ai decreti attuali della legge n. 273/1991 che ha istituito il Sistema Nazionale di Taratura (SNT). ACCREDIA attesta le capacità di misura e di taratura, le competenze metrologiche del Centro e la riferibilità delle tarature eseguite al campioni nazionali e internazionali delle unità di misura del Sistema Internazionale delle Unità (SI). Questo certificato non può essere riprodotto in modo parziale, salvo espressa autorizzazione scritta da parte del Centro.

Si riferisce a  
Referring to

- oggetto <i>item</i>	Flussimetro massico – Controllo strumento
- costruttore <i>manufacturer</i>	MKS - MKS
- modello <i>model</i>	GE50A013104RMV020 – PR4000BF2V2
- matricola <i>serial number</i>	022239841 – 018957352
- data di ricevimento oggetto <i>date of receipt of item</i>	25 gennaio 2021
- data delle misure <i>date of measurements</i>	9 febbraio 2021
- registro di laboratorio <i>laboratory reference</i>	R20-27

This certificate of calibration is issued in compliance with the accreditation LAT 159 granted according to decrees connected with Italian law No. 273/1991 which has established the National Calibration System. ACCREDIA attests the calibration and measurement capability, metrological competence of the Centre and the traceability of calibration results to the national and international standards of the International System of Units (SI). This certificate may not be partially reproduced, except with the prior written permission of the issuing Centre.

I risultati di misura riportati nel presente Certificato sono stati ottenuti applicando le procedure di taratura citate alla pagina seguente, dove sono specificati anche i campioni o gli strumenti che garantiscono la catena di riferibilità del Centro e i rispettivi certificati di taratura in corso di validità. Essi si riferiscono esclusivamente all'oggetto in taratura e sono validi nel momento e nelle condizioni di taratura, salvo diversamente specificato.

The measurement results reported in this Certificate were obtained following the calibration procedures given in the following page, where the reference standards or instruments are indicated which guarantee the traceability chain of the laboratory, and the related calibration certificates in the course of validity are indicated as well. They relate only to the calibrated item and they are valid for the time and conditions of calibration, unless otherwise specified.

Le incertezze di misura dichiarate in questo documento sono state determinate conformemente alla Guida ISO/IEC 98 e al documento EA-4/02. Solitamente sono espresse come incertezza estesa ottenuta moltiplicando l'incertezza tipo per il fattore di copertura  $k$  corrispondente ad un livello di fiducia di circa il 95 %. Normalmente tale fattore  $k$  vale 2.

The measurement uncertainties stated in this document have been determined according to the ISO/IEC Guide 98 and to EA-4/02. Usually, they have been estimated as expanded uncertainty obtained multiplying the standard uncertainty by the coverage factor  $k$  corresponding to a confidence level of about 95%. Normally, this factor  $k$  is 2.

Direzione tecnica  
(Approving Officer)

*Andrea Jona L.*



## **Annex n. 6096/B**



# Rapporto di Prova PD22-00604\_0



LAB N° 0080 L

## Prima pagina

### CLIENTE

Cliente NEW POWER ENERGY KFT

Indirizzo Kassai utca 9  
Budapest 1043

Progetto Default Project

Ordine n° 109/2022/C2/PD/Rev.2

Matrice ARIA: FLUSSI GASSOSI CONVOGLIATI

Prelevato presso Monte e valle sistema di filtrazione NPE

Prelevato da Ns. personale - Hendriksen, Bortoletti

### LABORATORIO

Head of Laboratory

Laboratorio

Indirizzo

Gianluca Martinati

SGS Italia S.p.A.

Via Campodoro, 25

Villafranca Padovana (PD) 35010

Telefono

+39 049 9050013

Fax

+39 049 9050065

Email

sgs.eco@sgs.com

Accettazione n°

PD22-00604

Pervenuto il

25/02/2022

Data inizio analisi

25/02/2022

Data fine analisi

28/02/2022

Data emissione

01/03/2022

### COMMENTI

Incertezza estesa di misura stimata al 95% di livello di confidenza e fattore di copertura k=2

### RIFERIMENTI

Enrico Costa  
Project Leader



**INDICE**

---

Prima pagina.....	1
Indice.....	2
Specifiche tecniche.....	3-5
Commenti operativi.....	6
Parametri di campo.....	7
Risultati.....	8
Legenda.....	9



## SPECIFICA TECNICA

**Punto di campionamento:**

Monte Identificativo presente

**Inizio campionamento:**

25/02/2022 10:30

**Fine campionamento:**

25/02/2022 11:00

**Requisiti sezione di campionamento e piattaforma di lavoro [UNI EN 15259:2008 sez. 6.2.1 e 8.2]**

Caratteristiche della piattaforma	Valore	Unità di misura	
Dimensione area di lavoro:	5,0	m <sup>2</sup>	
Altezza area di lavoro da terra:	1,0	m	
Caratteristiche del condotto	Valore	Unità di misura	
Geometria sezione:	Circolare	-	
Disposizione	Orizzontale	-	
Uscita diretta in atmosfera	SI	-	
Sezione condotto	0,005	m <sup>2</sup>	
Condizioni verificate in campo	Atteso	Misurato	Unità di misura
Tratto rettilineo a monte della sezione di campionamento (5 diametri idraulici):	40	105	cm
Tratto rettilineo a valle della sezione di campionamento ( 5 diametri idraulici):	40	110	cm
Spazio totale antistante ai punti di campionamento (Ø + tronchetto + 150 cm):	188	200	cm
Numero di assi da esplorare (Ø < 0,35 m 1 asse; Ø>0,35 2 o più assi):(#)	1	1	-
Flussi negativi misurati sulla sezione di campionamento (##):	NO	NO	-
Velocità minima riferita alla misura con Pitot (dp minimo 5 Pascal)(##):	5	8	Pa
Rapporto velocità massima e minima misurata sul piano (##):	≤ 3:1	1,00	-
Fattore correttivo sulla velocità:	0,995		

<sup>(#)</sup>Condizioni Vincolanti

**Premisurazioni e tarature in campo [UNI EN ISO 16911-1:2013 Annex A]**

Controllo funzionamento strumentazione	Esito test
Integrità Pitot - pre misura	SUPERATO
Integrità Pitot - post misura	SUPERATO
Leak check	SUPERATO
Test ostruzioni	SUPERATO
Verifica flussi nel condotto	Esito test
Variazione flusso picco/picco	POSITIVO
Ripetibilità su un punto del reticollo	POSITIVO
Angolo di swirl medio	4,0 °
Correzione necessaria	NO

*Nota: la correzione per l'angolo di swirl è necessaria solo se l'angolo è > 15°*
**Portata [UNI EN ISO 16911-1:2013 Annex A]      Unità di misura      Valore**

A 1	Punti traversi	Unità di misura	Valore
Temperatura		K	352,2
Pressione differenziale		Pa	7,7
Velocità		m/s	3,4



Punto di campionamento:	Valle	Identificativo presente
Inizio campionamento:	25/02/2022	11:00
Fine campionamento:	25/02/2022	11:30

**Requisiti sezione di campionamento e piattaforma di lavoro [UNI EN 15259:2008 sez. 6.2.1 e 8.2]**

Caratteristiche della piattaforma	Valore	Unità di misura	
Dimensione area di lavoro:	5,0	m <sup>2</sup>	
Altezza area di lavoro da terra:	1,0	m	
Caratteristiche del condotto	Valore	Unità di misura	
Geometria sezione:	Circolare	-	
Disposizione	Orizzontale	-	
Uscita diretta in atmosfera	Sì	-	
Sezione condotto	0,005	m <sup>2</sup>	
Condizioni verificate in campo	Atteso	Misurato	Unità di misura
Tratto rettilineo a monte della sezione di campionamento (5 diametri idraulici):	40	105	cm
Tratto rettilineo a valle della sezione di campionamento ( 5 diametri idraulici):	40	110	cm
Spazio totale antistante ai punti di campionamento ( $\varnothing$ + tronchetto + 150 cm):	188	200	cm
Numero di assi da esplorare ( $\varnothing < 0,35$ m 1 asse; $\varnothing > 0,35$ 2 o più assi):(#)	1	1	-
Flussi negativi misurati sulla sezione di campionamento (#):	NO	NO	-
Velocità minima riferita alla misura con Pitot (dp minimo 5 Pascal)(#):	5	5	Pa
Rapporto velocità massima e minima misurata sul piano (#):	≤ 3:1	1,00	-
Fattore correttivo sulla velocità:	0,995		

<sup>(#)</sup>Condizioni Vincolanti

**Premisurazioni e tarature in campo [UNI EN ISO 16911-1:2013 Annex A]**

Controllo funzionamento strumentazione	Esito test
Integrità Pitot - pre misura	SUPERATO
Integrità Pitot - post misura	SUPERATO
Leak check	SUPERATO
Test ostruzioni	SUPERATO
Verifica flussi nel condotto	Esito test
Variazione flusso picco/picco	POSITIVO
Ripetibilità su un punto del reticolino	POSITIVO
Angolo di swirl medio	3,0 °
Correzione necessaria	NO

*Nota: la correzione per l'angolo di swirl è necessaria solo se l'angolo è > 15°*

Portata [UNI EN ISO 16911-1:2013 Annex A]	Punti traversi	Unità di misura	Valore
A 1			
Temperatura		K	284,2
Pressione differenziale		Pa	5
Velocità		m/s	2,4

**ACCREDIA**  
L'ENTE ITALIANO DI ACCREDITAMENTO

LAB N° 0080 L

## SPECIFICA TECNICA

Metodo	Inizio	Fine	V Normalizzato 0°C 1013 hPa
UNI EN ISO 23210:2009	25/02/2022 11:00	25/02/2022 12:00	1,3470
UNI EN ISO 23210:2009	25/02/2022 12:15	25/02/2022 13:15	2,3380



## COMMENTI OPERATIVI

I valori di concentrazione sono normalizzati (101.3 kPa - 273 K) ed espressi su fumi secchi.

Sigla campione	Campione n°	Analisi	Commenti
ANALISI STRUMENTALE VALLE	PD22-00604.004	Blossido di carbonio	CO2: 148,6 ppm (misura eseguita con TSI 7545 IAQ-Calc)


**Parametri di campo**

	Campionato a	MONTE	VALLE
Inizio campionamento	25/02/2022 10:30	25/02/2022 11:00	
Fine campionamento	25/02/2022 11:00	25/02/2022 11:30	
Parametro	U.M.	Risultato	Risultato

**Parametri di campo**
**Portata [ UNI EN ISO 16911-1:2013 Annex A ]**

A Sezione camino	m <sup>2</sup>	0,005	0,005
A Portata fumi umidi	Nm <sup>3</sup> /h	47,00 ± 1,00	41,00 ± 1,00
A Portata fumi secchi	Nm <sup>3</sup> /h	36,00 ± 2,00	41,00 ± 2,00
A Temperatura	°C	79,00 ± 5,00	11,00 ± 4,00
A Velocità	m/s	3,360 ± 0,080	2,360 ± 0,060
A Pressione statica assoluta media	hPa	1015,00 ± 10,20	1015,00 ± 10,20
A Massa Volumica del gas	Kg/m <sup>3</sup>	0,94	1,2
A Fattore di taratura del tubo di Pitot tipo S	-	0,83	0,83

**Ossigeno (O<sub>2</sub>) [ UNI EN 14789:2017 ]**

A Ossigeno (O <sub>2</sub> )	%v/v	1,82 ± 0,11	20,9 ± 1,5
------------------------------	------	-------------	------------

**Ossigeno, Biossido di carbonio - metodo strumentale [ EPA 3A 2017 ]**

A Biossido di carbonio	%v/v	12,37 ± 0,40	<0,10
------------------------	------	--------------	-------

**Vapore acqueo [ UNI EN 14790:2017 ]**

A Volume campionario	Nm <sup>3</sup>	0,155	0,155
A Vapore acqueo raccolto	mL	39	<1
A Vapore acqueo (su base umida)	%v/v	23,90 ± 2,30	<2,0


**Risultati**

	Campionato a	MONTE	VALLE
Inizio campionamento	25/02/2022 10:30	25/02/2022 10:30	
Fine campionamento	25/02/2022 13:00	25/02/2022 13:00	
Parametro	U.M.	Risultato	Risultato

**Ossigeno (O<sub>2</sub>) [ UNI EN 14789:2017 ]**

A Ossigeno (O <sub>2</sub> )	%v/v	1,89 ± 0,12	20,9 ± 1,5
------------------------------	------	-------------	------------

**Biossido di carbonio - metodo strumentale [ EPA 3A 2017 ]**

A Biossido di carbonio	%v/v	12,33 ± 0,39	<0,10
------------------------	------	--------------	-------

**Biossido di zolfo [ UNI CEN/TS/17021:2017 ]**

*A Biossido di zolfo	mg/Nm <sup>3</sup>	<2,9	<2,9
----------------------	--------------------	------	------

**Ossidi di azoto (NOx) [ UNI EN 14792:2017 ]**

A Ossidi di azoto (NOx)	mg/Nm <sup>3</sup>	5400 ± 460	<2,1
-------------------------	--------------------	------------	------

**Composti Organici Volatili (COV) espressi come Carbonio Organico Totale [ UNI EN 12619:2013 ]**

A Composti Organici Volatili (COV) espressi come Carbonio Organico Totale	mg/Nm <sup>3</sup>	9,03 ± 0,90	1,73 ± 0,17
---	--------------------	-------------	-------------

**Monossido di Carbonio [ UNI EN 15058:2017 ]**

A Monossido di Carbonio	mg/Nm <sup>3</sup>	<1,3	<1,3
-------------------------	--------------------	------	------

	Campionato a	MONTE	VALLE
Inizio campionamento	25/02/2022 11:00	25/02/2022 12:15	
Fine campionamento	25/02/2022 12:00	25/02/2022 13:15	
Parametro	U.M.	Risultato	Risultato

**PM 2.5 - 10 [ UNI EN ISO 23210:2009 ]**

A PM 2,5	mg/Nm <sup>3</sup>	0,082 ± 0,023	<0,043
A PM 10	mg/Nm <sup>3</sup>	0,341 ± 0,097	0,064 ± 0,018



## LEGENDA

## NOTE

^	Eseguito presso laboratorio SGS esterno.	IS	Campione insufficiente per l'analisi.
^^	Eseguito presso laboratorio esterno.	LNR	Campione elencato ma non ricevuto.
RL	Limite di Rapportaggio	NA	Campione non analizzato per questo parametro
↑	Limite di rapportaggio innalzato	TBA	Parametro non ancora analizzato
↓	Limite di rapportaggio diminuito	†	Tempo massimo di conservazione superato

## NOTE RELATIVE ALL'ACCREDITAMENTO

- Prova non accreditata ACCREDIA.

Il presente Rapporto è emesso dalla Società in accordo con le Condizioni Generali SGS per i servizi di ispezione e controllo (copia disponibile su richiesta). Il rilascio di questo Rapporto non esonerà le parti negoziali dall'esercitare i diritti e dall'adempire alle obbligazioni derivanti dal negozio tra loro stipulato. Ogni patto contrario non è alla Società opponibile. La responsabilità della Società in base a questo Rapporto è limitata al caso di provata colpa grave ed in ogni caso ad un ammontare non superiore a dieci volte i diritti e le commissioni dovute. Eccetto accordi particolari, gli eventuali campioni, se presi, non saranno trattenuti dalla Società per più di un mese. I riscontri analitici ed i risultati delle elaborazioni si riferiscono esclusivamente alle condizioni operative in atto nel periodo in cui è stata effettuata la presente indagine.

Il Laboratorio declina ogni responsabilità sui dati forniti dal cliente che possono influenzare la validità dei risultati. Il presente Rapporto o copia dello stesso verrà conservato dalla Società per un periodo pari a 10 anni.

Il recupero ove previsto, è da intendersi compreso all'interno dei limiti di accettabilità specifici (70-130% per microinquinanti ORGANICI, 75-125% per microinquinanti INORGANICI). Se non diversamente indicato il risultato è da intendersi non corretto per il recupero ottenuto.

Se non diversamente specificato, valori di concentrazione rilevati inferiori ai Limite di Rapportaggio (RL) concorrono all'espressione delle somme e/o medie nella misura di 1/2 del Limite di Rapportaggio ( criterio "medium bound").

A=Prova eseguita presso la sede di SGS Italia SpA Via Campodoro 25 – 35010 Villafranca Padovana (PD) – ITALIA

B=Prova eseguita presso la sede di SGS Italia SpA Via Campodoro 23 – 35010 Villafranca Padovana (PD) – ITALIA

C=Prova eseguita presso la sede di SGS Italia SpA Quarta Strada Z.I. Macchiareddu - 09032 Assemini (CA) – ITALIA

D=Prova eseguita presso la sede di SGS Italia SpA C.da Spalla Città Giardino - 96010 Melilli (SR) – ITALIA

Il presente rapporto può essere riprodotto solamente per intero.

--- Fine del Rapporto di Prova ---